

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2019/11/10
Person Assigned	Kevin O'Donnell
Submitter Name	JIRA
Submission Date	2019/02/04

Correction Number	CP-1912
Log Summary:	Correct Retired Attributes in PS3.15 E
Name of Standard	PS 3.15 2019d
Rationale for Correction:	Based on PS3.6, the retirement status shown in PS3.15 Table E.1-1 needs to be either corrected or added for several attributes.
Correction Wording:	

Modify PS3.15 to update 3rd and 4th columns of E.1-1 as shown

E.1.1 De-Identifier

...

Table E.1-1. Application Level Confidentiality Profile Attributes

Attribute Name	Tag	Retd. (from PS3.6)	In Std. Comp. IOD (from PS3.3)	...
...				
Medical Record Locator	(0010,1090)	<u>Y</u> N	N	...
..				
Patient's Insurance Plan Code Sequence	(0010,0050)	<u>N</u>	<u>Y</u>	...
...				
Patient's Primary Language Code Sequence	(0010,0101)	<u>N</u>	<u>Y</u>	...
Patient's Primary Language Modifier Code Sequence	(0010,0102)	<u>N</u>	<u>Y</u>	...
...				
Related Frame of Reference UID	(3006,00C2)	<u>Y</u> N	<u>N</u> Y	...
...				